



NABL

**National Accreditation Board for
Testing and Calibration Laboratories**
Department of Science & Technology, India

CERTIFICATE OF ACCREDITATION

UNIVERSAL CABLES LIMITED, EHV CABLE LABORATORY

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

P.O. Birla Vikas, Satna

in the field of

ELECTRICAL TESTING

(You may also visit NABL website www.nabl-india.org to view the scope of accreditation)

Certificate Number T-2043

Issue Date 19/08/2011

Valid Until 18/08/2013

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the additional requirements of NABL.

Signed for and on behalf of NABL

Spikanth. R
Convenor

Anil Relia
Director

Dr. T. Ramasami
Chairman



NABL

Department of Science & Technology, India

SCOPE OF ACCREDITATION

Laboratory	Universal Cables Limited, EHV Cable Laboratory, P.O. Birla Vikas, Satna, Madhya Pradesh		
Accreditation Standard	ISO/IEC 17025: 2005		
Field	Electrical Testing	Issue Date	19.08.2011
Certificate Number	T-2043	Valid Until	18.08.2013
Last Amended on	-	Page	1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
I.	EHV CABLES	Partial Discharge Test	IEC 60840 IEC 62067 IS 7098 (Part III)	0 to 500 pc/ 1pc
		High Voltage Test	IEC 60840 IEC 62067 IS 7098 (Part III)	0 to 650 kV/ 1kV
		Tan Delta Test	IEC 60840 IEC 62067 IS 7098 (Part III)	0 to 300 kV/ 1x10 ⁵
		Impulse Voltage Test	IEC 60840 IEC 62067 IS 7098 (Part III)	0 to 2400 kV/ 10kV

-X-X-X-X-X-X-X-X-X-X-X-

Convenor